Features

- 1.8"x3.2"x1.2", encapsulated module
- 40W power from -40°C up to +65°C ambient
- Operating temp. up to +85°C with derating

4 kVac/1min reinforced isolation

- Regulated Converter
- 2MOPP medical certified, B and BF compliant
 5000m (medical/ITE) operating altitude
- Class B EMC filter built-in

Description

The ultra-compact encapsulated industrial + household + medical grade AC/DC converter series RACM40-K delivers 40 watts of output power from -40°C to +65°C with natural air convection only, and up to +85°C with derating or forced air cooling. With a clear focus on extended thermal performance for systems where space is limited, these 1.8" x 3.2" compact modules are designed to gain highest overall efficiency levels over the full output load range from universal AC inputs. The RACM40-K has ANSI/ AAMI/IEC 60601-1 medical safety and EN 60601-1-2 medical EMC certifications and offers 4kVac/1 min isolation, 2MOPP, and is designed to meet B and BF requirements. It is additionally certified (CB Report) to IEC/EN 62368-1; IEC61010 and IEC61558-1/-2-16 for industrial applications and IEC/EN 60335-1 for household appliances. The robust built-in class B EMC filter has sufficient margin to allow either Class II or Class I PELV with grounded output installations. The mechanically rugged construction with fully potted encapsulation, 1,6mm pins and additional threaded inserts gives the series enhanced stability against shock and vibrations.

Selection Guide					
Part Number	Input Voltage Range [VAC]	Output Voltage [VDC]	Output Current [mA]	Efficiency typ. ⁽¹⁾ [%]	Max. Output Power [W]
RACM40-05SK-T	80-264	5	6000	87	30
RACM40-12SK-T	80-264	12	3334	90	40
RACM40-15SK-T	80-264	15	2667	90	40
RACM40-24SK-T	80-264	24	1667	90	40
RACM40-48SK-T	80-264	48	833	90	40

Notes:

Note1: Efficiency is tested at +25°C with constant resistant mode at full load and 230VAC

Selection Guide	e (on request MO	Q ≥1008pcs)		
Part Number	Input Voltage Range [VAC]	Output Voltage [VDC]	Output Current [mA]	Efficiency typ. ⁽¹⁾ [%]	Max. Output Power [W]
RACM40-18SK-T	80-264	18	2222	90	40
RACM40-36SK-T	80-264	36	1111	90	40

RECOM AC/DC Converter

RACM40-K

40 Watt 1.8" x 3.2" Single Output



IEC/EN62368-1 (pending) ANSI/AAMI ES60601-1 certified CSA/CAN-C22.2 No. 60601-1:14 certified IEC/EN60601-1 certified IEC/EN60335-1 (pending) IEC/EN61010-1 (pending) EN62233 (pending) IEC/EN61558-1 (pending) IEC/EN61558-2-16 (pending) EN55032/35 compliant IEC/EN60601-1-2 compliant CB Report (pending)

Model Numbering



RACM40-K

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)

Series

Parameter		Cond	dition	Min.	Тур.	Max.
Nominal Input Voltage		60)Hz	100VAC		
Nominal liput voltage		50)Hz			240VAC
Operating Range ⁽²⁾	47-63Hz			80VAC		264VAC
		C	DC	120VDC		370VDC
Input Current			5VAC			1000mA
		230	DVAC			500mA
Inrush Current	cold	start	115VAC			15A
	cold start		230VAC			30A
No load Power Consumption		230	OVAC		100mW	
	RACM40		input power max. 0.5W	0.3W		
ErP Standby Mode Conformity (Maximum output power available for stated maximum input power)	115VAC -	RACM40	input power max. 1.0W	0.7W		
		RACM40	input power max. 0.5W	0.27W		
	230VAC	RACM40 input power max. 1.0W		0.65W		
Input Frequency Range				47Hz		63Hz
Minimum Load				0%		
Power Factor		115	5VAC	0.6		
POWEI Faciol	230VAC		0.5			
Start-up Time				160ms		
Rise Time				70ms		
Hold-up Time	115VAC		16ms			
	230VAC		60ms			
Internal Operating Frequency		100% load a	at nominal Vin		100kHz	
Output Ripple and Noise (3)	20MHz BW 5Vout		5Vout			80mVp-p
		2010112 BW others				1% of Vout


Notes:


Note2:

: The products were submitted for safety files at AC-Input operation

Note3: Measurements are made with a 0.1µF MLCC & 10µF E-cap in parallel across output. (low ESR)

Efficiency vs. Load






www.recom-power.com

RACM40-K Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)



PROTECTIONS				
Parameter	Туре	Value		
Internal Input Fuse		T3.15A, slow blow type		
Short Circuit Protection (SCP)	below 100mΩ	hiccup, auto recovery		
Over Voltage Protection (OVP)		105% - 120% of nom. Vout, hiccup mode		
Output reverse Voltage Protection	overrun rate of nominal output	107% - 145% of nom. Vout, hiccup mode		

continued on next page

Downloaded from Arrow.com.

RACM40-K **Series**

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)


Parameter	Туре	e/Condition	Value
Over Current Protection (OCP)			130% - 180% of nom. lout, hiccup mode
Thermal Shutdown	measured on TC point refe	to "Dimension Drawing (mm)"	+130°C typ.
Over Voltage Category (OVC)			OVCII
Class of Equipment			Class II
Isolation Voltage (safety certified) (5)	I/P to O/P	1 minute	4kVAC
Isolation Resistance	I/P to O/P	I/P to O/P, Viso= 500VDC	1GΩ min.
Isolation Capacitance	I/P to O/P	I/P to O/P, 100KHz/0.1V	100pF max.
Leakage Current			1.5mA max.
Insulation Grade			reinforced
	tes:		
	Note5: For repeat Hi-Pot testing, re	duce the time and/or the test voltage	

ENVIRONMENTAL Parameter Condition Value @ natural convection 0.1m/s -40°C to +60/65°C without derating **Operating Temperature Range** -40°C to +85°C (refer to "Derating Graph") with derating 100°C Max. Case Temperature ±0.02%/K **Temperature Coefficient** Thermal Impedance 6.3K/W Operating Altitude (6) according to 62368-1/61010 and 60601-1 5000m **Operating Humidity** non-condensing 20% - 95% RH max **Pollution Degree** PD2 Vibration according to MIL-STD-202G 10-500Hz, 2G 10min./1cycle, period 60min. along x,y,z axes +25°C >1006 x 103 hours MTBF according to MIL-HDBK-217F, G.B. +40°C >790 x 103 hours Design Lifetime 230VAC/60Hz and full load +40°C >98 x 103 hours

Notes:

Note6: Recognized by safety agency for safe operation up to 5000m. High altitude operation may impact the performance and lifetime. Please contact RECOM tech support for advice

Derating Graph



www.recom-power.com

65

85

RACM40-K Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)

SAFETY AND CERTIFICATIONS		
Certificate Type (Safety)	Report / File Number	Standard
Medical electrical equipment Part 1: General requirements for basic safety and essential performance		ANSI/AAMI ES60601-1:2005 + A2:2010/2012 CAN/CSA-C22.2 No. 60601-1:14, 3rd Edition
Medical electrical equipment Part 1: General requirements for basic safety and essential performance	E511305-D1001-1/A0/C0-UL	IEC60601-1:2005, 3rd Edition + AM1:2012 EN60601-1:2006 + A1:2013
Audio/Video, information and communication technology equipment - Safety requirements (CB Scheme)	pending	IEC62368-1:2014 2nd Edition
Audio/Video, information and communication technology equipment - Safety requirements (LVD)	pending	EN62368-1:2014 + A11:2017
Household and similar electrical appliances - Safety - Part 1: General requirements	pending	IEC60335-1:2010 5th Edition + C1:2016
Household and similar electrical appliances – Safety – Part 1: General requirements (LVD)	pending	EN60335-1:2012 + A14:2019
Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements (CB Scheme)	pending	IEC61010-1:2010+A1:2016, 3rd Edition
Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements	pending	EN61010-1:2010+A1:2019
Measurement methods for electromagnetic fields of household appliances and similar apparatus with regard to human exposure	pending	EN62233:2008
Safety of power transformers, power supplies, reactors & similar products for supply voltages up to 1100V (CB Scheme)	pending	IEC61558-1:2005 2nd Edition + A1:2009
Safety of power transformers, power supplies, reactors & similar products for supply voltages up to 1100V	pending	EN61558-1:2005 + A1:2009
Safety of power transformers, power supplies, reactors & similar products for supply voltages up to 1100 V Part 2: Particular requirements (CB Scheme)	pending	IEC61558-2-16:2009 1st Edition + A1:2013
Safety of power transformers, power supplies, reactors & similar products for supply voltages up to 1100 V Part 2: Particular requirements	pending	EN61558-2-16:2009 + A1:2013
RoHS2		RoHS 2011/65/EU + AM2015/863
EMC Compliance (Medical)	Condition	Standard / Criterion
Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Re- quirements and tests 4th Ed.	4789293779	EN60601-1-2:2015
ESD Electrostatic discharge immunity test	Air ±2, 4, 8, 15kV; Contact ±8kV	IEC61000-4-2:2008 , Criteria A EN61000-4-2:2009, Criteria A
Radiated, radio-frequency, electromagnetic field immunity test	9V/m (710, 745, 780, 5240, 5500, 5785MHz) 10V/m (80-2700MHz) 27V/m (385MHz) 28V/m (450, 810, 870, 930, 1720, 1845, 1970, 2450MHz)	IEC/EN61000-4-3:2006 + A2:2010, Criteria A
Fast Transient and Burst Immunity	AC Por:t L, N, L-N ±2kV	IEC/EN61000-4-4:2012, Criteria A
Surge Immunity	AC Port L-N: ±0.5, 1, 2kV L-PE, N-PE: ±0.5, 1, 2, 4kV	IEC/EN61000-4-5:2014, Criteria B
Immunity to conducted disturbances, induced by radio-frequency fields	AC Port: 3Vrms (0.15-80MHz) 6Vrms (IMS Band)	IEC61000-4-6:2013, Criteria A EN61000-4-6:2014, Criteria A
Power Magnetic Field Immunity	30A/m	IEC61000-4-8:2009, Criteria A EN61000-4-8:2010, Criteria A
Voltage Dips and Interruptions	Voltage Dips 30% Voltage Dips 100% (0.5P) Voltage Dips 100% (1.0P) Voltage Interruptions 100%	IEC/EN61004-11:2004, Criteria A IEC/EN61004-11:2004, Criteria A IEC/EN61004-11:2004, Criteria A IEC/EN61004-11:2004, Criteria B

continued on next page

Downloaded from Arrow.com.

RACM40-K

Series

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)

EMC Compliance (Industrial)	Condition	Standard / Criterion
Electromagnetic compatibility of multimedia equipment – Emission Requirements		EN55032:2015
Electromagnetic compatibility of multimedia equipment – Immunity requirements	LCS200616044BE	EN55035:2017
ESD Electrostatic discharge immunity test	Air ±2, 4, 8kV; Contact ±2, 8kV	IEC61000-4-2:2008 , Criteria A EN61000-4-2:2009, Criteria A
Radiated, radio-frequency, electromagnetic field immunity test	3V/m (4800-1000MHz, 1800, 2600, 3500, 5000MHz)	IEC/EN61000-4-3:2006 + A2:2010, Criteria A
Fast Transient and Burst Immunity	AC Port: L, N, L-N ±1kV	IEC/EN61000-4-4:2012, Criteria B
Surge Immunity	AC Port: L-N: ±1kV	IEC/EN61000-4-5:2014, Criteria B
Immunity to conducted disturbances, induced by radio-frequency fields	AC Port: 3Vrms (0.15-80MHz) 3Vrms (10-30MHz) 1Vrms (30-80MHz)	IEC61000-4-6:2013, Criteria A EN61000-4-6:2014, Criteria A
Power Magnetic Field Immunity	1A/m	IEC61000-4-8:2009, Criteria A EN61000-4-8:2010, Criteria A
Voltage Dips and Interruptions	Voltage Dips 30% Voltage Dips 100% Voltage Interruptions 100%	IEC/EN61004-11:2004, Criteria C IEC/EN61004-11:2004, Criteria B IEC/EN61004-11:2004, Criteria C


EMC Compliance (Low voltage power supply)	Condition	Standard / Criterion
Low voltage power supplies, d.c. output Part 3: Electromagnetic compatibility (EMC)	LCS200616049BE	IEC/EN61204-3:2018
ESD Electrostatic discharge immunity test	Air ±2, 4, 8kV;	IEC61000-4-2:2008, Criteria A
	Contact ±2, 8kV	EN61000-4-2:2009, Criteria A
	10V/m (80-1000MHz)	
Radiated, radio-frequency, electromagnetic field immunity test	3V/m (1400-2000MHz)	IEC/EN61000-4-3:2006 + A2:2010, Criteria A
	1V/m (2000-2700MHz)	
Fast Transient and Burst Immunity	AC Port: L, N, L-N ±2kV	IEC/EN61000-4-4:2012, Criteria B
Surge Immunity	AC Port: L-N: ±1kV	IEC/EN61000-4-5:2014, Criteria B
Immunity to conducted disturbances, induced by radio-frequency fields	AC Port: 10Vrms (0.15-80MHz)	IEC61000-4-6:2013, Criteria A
		EN61000-4-6:2014, Criteria A
Power Magnetic Field Immunity	30A/m	IEC61000-4-8:2009, Criteria A
	SUAVIII	EN61000-4-8:2010, Criteria A
	Voltage Dips 20, 30,60%	IEC/EN61004-11:2004, Criteria C
Veltage Dipe and Interruptions	Voltage Dips 100% (0.5P)	IEC/EN61004-11:2004, Criteria B
Voltage Dips and Interruptions	Voltage Dips 100% (1.0P)	IEC/EN61004-11:2004, Criteria B
	Voltage Interruptions 100%	IEC/EN61004-11:2004, Criteria C
Limits of Voltage Fluctuations & Flicker		EN61000-3-3:2013
Limitations on the amount of electromagnetic interference allowed from digital and electronic devices		FCC 47 CFR Part 15 Subpart B, Class B
Limitations on the amount of electromagnetic interference allowed from digital and electronic devices, industrial, scientific, and medical equipment		FCC 47 CFR Part 18

DIMENSION AND PHYSICAL CHARACTERISTICS				
Parameter	Туре	Value		
	PCB	FR4, (UL94 V-0)		
Material	potting	PU, (UL94 V-0)		
	baseplate	plastic, (UL94V-0)		
Dimension (LxWxH)		83.23 x 46.43 x 30.40mm		
Weight		185g typ.		

continued on next page

RACM40-K **Series**

Specifications (measured @ Ta= 25°C, nom. Vin, full load and after warm-up unless otherwise stated)



The product information and specifications may be subject to changes even without prior written notice. The product has been designed for various applications; its suitability lies in the responsibility of each customer. The products are not authorized for use in safety-critical applications without RECOM's explicit written consent. A safety-critical application is an application where a failure may reasonably be expected to endanger or cause loss of life, inflict bodily harm or damage property. The applicant shall indemnify and hold harmless RECOM, its affiliated companies and its representatives against any damage claims in connection with the unauthorized use of RECOM products in such safety-critical applications.

Downloaded from Arrow.com.